

Visy Pulp and Paper Tumut CEMS - Correction Log

8/04/2020

Reporting Period: 1/03/2020 - 1/04/2020

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
10/03/20 14:36	10/03/20 14:54	Instrument Error	Normal (Steady State)	2 Monthly calibration of Main stack A	37.92

Main Stack 2

Monitoring Location No: 22
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 2 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
09/03/20 12:48	09/03/20 13:00	Instrument Error	Normal (Steady State)	Calibration of Both ESP duct opacities and also stack opacity	41.19

TRS (as H2S)		Period: 60 Minutes	Limit: 3.60 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading

22/03/20 02:00	22/03/20 03:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	5.74
29/03/20 02:00	29/03/20 03:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	5.79

Authorised By:

Uday Bhagwat
Pulp Mill Manager



Johan Stoltz
General Manager

